भारत सरकार/Government of India

स्वास्थ्यऔर परिवारकल्याणमंत्रालय/ Ministry of Health and Family Welfare प्रधानमंत्री स्वास्थ्य सुरक्षा योजना/PMSSY

अखिलभारतीयआयुर्विज्ञानसंस्थान/All India Institute of Medical Sciences www.aiimsmangalagiri.edu.in मंगलिगिरि, आंध्रप्रदेश/Mangalagiri, Andhra Pradesh

Ref: AIIMS/MG/MS/09/Procurement/Non DPR/High End ICU Beds

FTS: 3415 Date: 13-12-2023

Call for Objection

Subject: Inviting comments/objection, if any before declaring proprietary article for procurement of "**High End ICU Beds**" for O/o. of MS (for IPD) AIIMS Mangalagiri.

O/o. of MS, AIIMS Mangalaigiri has to procure "High End ICU Beds" through Proprietary Article basis.

The proposal submitted by M/s. Hanuman Health Care, New Delhi who is authorized distributor of the sole manufactures M/s. Famed Zyweic Sp.Zo.o, Fabryczna Poland of this item along with Proprietary Article Certificate are attached & uploaded on Institute website.

The above documents are being uploaded for open information to submit objections, comments if any from any manufacturer/supplier before declaring proprietary article of the said equipment/items to be procured, within 10 days from the date of issuance/uploading of the notification.

The objection should be raised in the technical compliance sheet as enclosed in Annexure -I, if any Firm claiming suitability of their product with respect to specification mentioned.

The comments should be sent to the office of Procurement Cell, 2nd floor, Admin cum Library building at AIIMS Mangalagiri in a sealed envelope with above reference on or before 23/12/2023 up to 05:00 PM from the date of uploading on institutional website, failing which it will be presumed that any other manufacture/vendor is having no comment to offer and case will be decided on merits.

-sd-F I/C Procurement AIIMS Mangalagiri

CP-14



P-3 FORM

(to be attached with P-2 form for Proprietary items)

AIIMS Mangalagiri

PROPRIETORY ARTICLE CERTIFICATE

It is certified that the item High end ICU Beds required in the P-2 form should be purchased from M/s Hanuman Health Care ,New Delhi .To the best of my knowledge M/s Famed Zywise Sp.Zo.o, Fabryczna 1 is the sole manufacturer of the said item.

The product with specifications attached and features in the bed are not manufactured by any other manufacturer.

(Sign of Indenter)

Dated

Designation

Department

Recommend at on: 2 6 10 2

Deputy Medical Sciences
All India Institute of Medical Sciences
Mangalagin, Andhra Pradesh
Mangalagin, Andhra Pradesh

डॉ. विनीत थॉमस अवाहम / Dr. Vineet Thomas Abraham

चिकित्सा अधीयक / Medical Superintendent अधिक भारतीय आयुर्धियान संरक्षान, मंगलगिरि (आ.प्र.)»

Signature of Head of Department/SectionAll India Institute of Medical Sciences, Mangalagiri (A.P.)

N.B.: The indenter before recording the above certificate should satisfy himself that the article is genuinely of proprietary nature manufactured under patent laws.



Date: 13.10.2023

PROPRIETARY CERTIFICATE AND LETTER OF CONFIRMATION

Dear Sir,

This is to confirm that ICU Bed model LE-13 with the feature of mounting on three cylindrical columns allowing use of C-Arm with the bed and ease of cleaning, facility to insert X-ray Cassette tray through a tunnel under back, seat and thigh section for performing X-ray on the bed without moving the patient, facility of electrical lateral tilt adjustment to allow conducting positional therapy and ease of patient handling, facility to achieve preprogrammed cardiological chair, fowler position, electric CPR, examination position, anti shock position by using single button on controller is a proprietary product of Famed Żyweic Sp.z.o.o Poland.

We also confirm that the said product with all its above published specifications and features in same bed is not manufactured by any other manufacturer in the world.

We are pleased to confirm that M/s Hanuman Healthcare, having its registered office at SF, D-71, Malviya Nagar, New Delhi - 110 017, India is authorized distributor for our Range of products in India

Thanking you,

Yours faithfully,

FAMED ŻYWIEC Sp. z o.o. 34-300 Żywiec, ul. Fabryczna 1 Reg. 146369568 NIP 5272685925 tel. 33 86 66 200 Vice President, CFO

Pawel Tomala

Signed:

Paweł Tomala Vice President, CFO Famed Żywiec Sp. z o. o.

Duly authorized to sign this Authorization for and on behalf of: Famed Żywiec Sp. z o. o

Famed Żywiec Sp. z o.o. ul Fabryczna 1, 34 300 Zywiec

t. +48 33 856 62 00 f. +48 33 475 58 90 e. export@famed.com Registered in District Court in Bielsko Biala VIII Commercial Division of the National Court Register: KRS number 0000437709 VAT ID-PL5222685925 REGON: 146369568, BDO: 000005595, Amount of initial capital: 31: 100.750,00 PLN:





CP-8

HANUMAN Healthcare India

Dated: 25th October, 2023

Delhi I Bombay I Lucknow I Guwahati I Cochin I Jammu I Bhubaneswar I Kathmandu E-mail: info@hanumanhealthcare.com Web Site: www.hanumanhealthcare.com

HH/23-24/1100

To

The Medical Superintendent All India Institute of Medical Sciences

Mangalagiri, Andhra Pradesh 522503

Sub: Your requirement for ICU Beds with lateral tilt Facility

Dear Sir,

Please refer to the above-mentioned subject; we are pleased to quote M/s. Famed Žywiec Sp. z o.o., Poland make ICU Bed Model: LE-13 with following features:

S. No.	Description	Qty.	Unit Price INR
1.	ICU bed Model LE-13	1	1
4.0	Features:		
	Equipped with Plastic head/foot boards, with Control system with scale measuring patient's weight Mounted on Three cylindrical columns to allow use of C-Arm		
	 Controlling bed functions via hand controller and central control panel (plus additional buttons: cardiologic chair, antishock, examination positions, Fowler position, electric CPR), 		
	Blocking of electric functions via central control panel, central switch or electric functions, lateral tilts, inbuilt battery Back up		
	a	15.5	1
	- Manual CPR function in back and thigh rest,		139
	- Auto contour function,	100	The state of the s
	- Pallet filled in Radiolucent HPL,		1
	Pattet filled in Radiotucent Fire, Facility to insert X-ray Cassette tray under back and thigh segment		
	- Facility to insert x-ray Cassette tray direct start	1	ARREST PER
	Chasis with ABS covers, Central brake system with steering, castors 150 mm,	1	
		1	
	- Pallet extension,	1	1 -1 -0 -1
	- Linen Shelf for bed clothes,	1	
	- 4 bumpers in corners,		
	- Rails for additional accessories,		
	- 4 hooks for fluids.		
	 4 hooks for fulds. Adjustments: Hi-low, lateral tilts, back, thigh rests, Trend/Rev. Trend adjustable by electric actuators, shank rest adjustable via gas spring. 		
	- Max. load 250 kg		
	Price of ICU bed in INR		
	Add: GST@18%		
	Total Price including GST		

We hope you will find the above in line to your requirement

Thanking you,

Yours Sincerely,

For Hanuman Healthbare India.

Kamat Kumar (Sr. Product Specialist)

SPECIFICATIONS

Specifications for I.C.U. Bed

Complete mattress base should be x-ray translucent, made of high pressure laminate for x-ray & ease of clearing / disinfection.

The bed should be mounted on three cylindrical columns allowing use of C-Arm with the bed and ease of cleaning to prevent infections in ICU

Should have the facility to insert X-ray Cassette tray through a tunnel under back, seat and thigh section for ease of performing patient X-ray on the bed without moving the patient.

ICU bed should have Integrated weighing scale for patient weight measurement upto 250 kg with accuracy of

Should provide low bed position upto 485 mm to enable patients get into and out of bed easily

Should have facility of double auto-regression function upto 20 cm to reduce the risk of spinal injury and bedsores

It should have inbuilt facility of lateral tilt adjustment operated by remote for conducting positional therapy, for comfort and safety of patients with injuries of Hip, Spine etc & for convenience of staff in patient handling.

8. It should have inbuilt facility of bed length extension for large patients upto 20 cm. Bed should have inbuilt retractable linen shelf

Electric functions should be operated by hand controller, central control panel and integrated controls in side

10. Bed should be equipped with split side rail with integrated feather touch controls for all movements on both sides of rail. Side rails should have switch for nurse call and under bed light for patient safety

11. It should be equipped with a central control panel with facility of locking the functions to prevent accidental change of position by patient / attendant.

12. central control panel should a LCD touch screen to select the adjustment and display all movements and the angle of adjustment etc.

13. Bed should have the facility to achieve pre-programmed position cardiological chair, fowler position, electric CPR, examination position, anti-shock position by using single button on controller.

14. Should have a manual quick release mechanism for back section and thigh rest section for emergency CPR
15. Should be equipped with large castors (diameter 150 mm or more) with central braking and steering facility.

16. Bed should have inbuilt battery backup to allow adjustments during transportation and power failure

17. Mattress of the Bed should be made up of High quality fire retardant, medical grade foam with Anti-Microbial agent incorporated into all components for Prohibiting growth of bacteria & fungi. Mattress should be original OEM product of the bed manufacturer, relevant certificate & catalogue should also be attached with certifications for the mattress material.

18. Base frame should be covered with plastic shrouds for ease of cleaning and disinfection

19. Should have bumpers at all four corners and place for fixing accessories

The firm/ foreign technology partner Should comply with safety and quality standards for medical equipment – European CE, MDD/93/42/EEC, ISO 13485 and ISO 14001

21. The firm/ foreign technology partner product Should comply with environment protection standard IP54 and protection class I against electrocution

22. Should have stepless adjustment for the following :-

Height: min. 485 mm or less (electrically)

Max. 880 mm or better (electrically)

Back section: 0-70 degrees or better (electrically)

Thigh Section: 0-40 degrees or better (electrically) Thigh Section

Shank Rest Adjustment : 22 degrees or better (pneumatically)
Trendelenburg : 15 degrees or better (electrically) e)

f)

Anti-trendelenburg : 13 degrees or better (Electrically)

Lateral Tilt : 0-25 degrees or better (Electrically)

Cardiological chair, fowler position, electric CPR, examination position, anti-shock position

Electrically operated by using single button on central controller.

Back Rest segment Autoregress: 120 mm

Thigh Rest Segment Autoregress: 90mm

23. Dimensions of bed

RELEEFE

2280 mm or better a) Length

950 mm or better b)

Mattress Size : 840 x 2060mm c)

250 kg or more Permissible load capacity

24. Each bed should be supplied complete with:

Bed Ends : 01 pair

Split Side Rails with controls :

IV Rod: 01 No.

Mattress: 01 No.

Objection should be submitted in following format:

S. no	Item specification as given	Specification offered by firm	Deviation if any	Remarks