



Date: **27/02/2024**

NIQ No: AIIMS/MG/Procurement/06/Serology & Immunology, Consumables & Reagents/Clinical Microbiology.

Sub: –“NIQ for Supply of Antiphospholipid Antibody Elisa Kit ” at AIIMS Mangalagiri.

### **Notice Inviting Quotations**

On behalf of The Director, AIIMS Mangalagiri Quotations are invited with price bid and other documents from eligible Manufacturers/ Firms/ Companies/ Authorized Agents/ Distributors/ Dealers/ Supplier Agencies for “**Supply of Antiphospholipid Antibody Elisa Kit**” at **AIIMS Mangalagiri**, as per specified terms and conditions.

#### **I. Schedule of Requirement:-**

S.No	Description of items	Required Quantity
1.	Antiphospholipid Antibody Elisa Kit (1 Kit = 96 Tests)	03 Kits (3 Kits =288 Tests)

#### **II. Specifications of Antiphospholipid Antibody Elisa Kit:**

1. ELISA Kit should able to detect IgG and IgM autoantibodies against Cardiolipin, Phosphatidyl serine, Phosphatidyl inositol, Phosphatidic acid and Beta-2-Glycoprotein I in human serum or plasma quantitatively.
2. Test should be performed in Human Serum, Plasma or Whole Blood.
3. Sensitivity should be  $\geq 98\%$ , Specificity should be  $\geq 98\%$ .
4. Kit size should be - 96 wells.
5. Individual Wells should be taken out to perform a smaller number of tests or to avoid well wastage.
6. Kit should comprise Calibrators, Reactive and Non-reactive Controls and all necessary reagents to perform the test.
7. Cold chain indicator to be provided with the kits.
8. A cumulative time/temperature indicator prequalified by WHO which indicates the exposure to temperature in the range of 2-8° C is to be provided on every pack of kits.
9. The supplier should ensure maintenance of cold chain during transportation of kits at 2° to 8° C.
10. Original kit literature in English/Hindi (not photocopy) provided with each kit in English mentioning the principle, components, methodologies, validity, criteria, performance characteristics, bio-safety, limitations of assay, storage condition, manufacturing and expiry dates and methods of disposal provided.
11. Kit should be compatible with all common ELISA readers and washers.
12. Kit should be approved from the statutory in its country of origin.
13. Kit should be CE-IVD approved.
14. Manufacture shall submit all necessary certifications, license and test reports to the buyer along with supplies.
15. Kit should have shelf life of 18-24 months from the date of manufacture.
16. Supplier should agree to provide advance sample of the kit for buyer's approval before commencement of bulk supply.
17. Kit should provide availability of test report/performance evaluation report from Central Govt/ NABL/ILAC accredited lab as well as in house test report to prove the conformity to the declared specification.

### **III. Terms and Conditions**

1. The sealed quotation must have valid GST number.
2. Rates quoted must include all types of taxes and other possible expenses. No other charges will be considered.
3. The payments will be made against delivery item with valid invoice and E-way bill through electronically viz RTGS/NEFT. The following information also to be mentioned in the quotation.
  - a. Name of the beneficiary:
  - b. Account No. of the beneficiary:
  - c. IFSC of the bank/branch:

**d). Delivery period:** Delivery period of the items shall be **30 days** from the issue of supply order. In the event of any deviation, the supply order may be cancelled.

**e)** The quotations will be **opened on 08/03/2024 03:00 PM** at Admin & Library Building, 2<sup>nd</sup> Floor, Room no.224 AIIMS Mangalagiri.

**f).** Vendors requested to attend at time of opening of bid.

**g) Penalty:** Penalty @ **0.50% per week** will levied for late delivery on supply order value of the item and will be deducted from the payment Bill. The maximum penalty will be **10 %** of total value against any of Purchase/Supply Order total value.

**h) Warranty Period:** Should be clearly mentioned in the quotation as per the OEM or company norms.

**i). Price Quotation:** price cannot exceed Rs.2.5 Lakhs.

**j). Price Validity:** The quoted prices should remain valid for a period of **180 days** from date of opening of NIQ.

**k).**The outer envelope should be super scripted as “**Supply of Antiphospholipid Antibody Elisa Kit**” at AIIMS Mangalagiri.

**IV.** The duly sealed and super scribed Quotations should reach **Procurement Cell, Admin & Library Building, 2<sup>nd</sup> Floor, Room no. 236, AIIMS Mangalagiri, Guntur-522503 by 03:00 PM on 07/03/2024.**

**Sd/-**  
**F I/c (Procurement)**  
**For and on behalf of Director**  
**AIIMS, Mangalagiri**