



अखिल भारतीय आयुर्विज्ञान संस्थान
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
मंगलगिरी, गुंटूर (आंध्र प्रदेश) 522503
Mangalagiri, Guntur District (Andhra Pradesh) – 522503
Website: <https://www.aiismangalagiri.edu.in/>

Date: 26/02/2025

WMS/NIQ No: AIIMS/MG/Procurement/06/Transport Monitor /Department of Cardio Vascular & Thoracic Surgery.
Sub: -“NIQ for Supply of Transport Monitor” at AIIMS Mangalagiri.

Notice Inviting Quotations

On behalf of The Executive Director, AIIMS Mangalagiri Quotations are invited with price bid and other documents from eligible Manufacturers/ Firms/ Companies/ Authorized Agents/ Distributors/ Dealers/ Supplier Agencies for **“Supply of Transport Monitor” at AIIMS Mangalagiri**, as per specified terms and conditions.

1. Schedule of Requirement: -

S.No	Description of Items	Required Quantity
1.	Transport Monitor	2 No's

II. Technical Specifications

The monitor should have:

1. High – resolution colour TFT display of minimum 8" or more
2. It should be rugged and sturdy for transport use.
3. Should be able to monitor ECG, NIBP, SpO2., Two IBP Temperature and Respiration
4. Plethysmograph with perfusion indicator
5. Monitor should be at least three channels
6. 24 Hrs. graphical / tabular trends
7. NIBP trends memory should be at least 50 readings (tabular)
8. Suitable for Adult / paediatric/neonate.
9. Selectable Arrhythmia detection
10. Should have inbuilt two channel recorder
11. Must have Graded and Colour coded alarms
12. User selectable screen formats and user – friendly menu driven functions.
13. Battery backup for at least 3 Hrs.
14. Should be supplied with:
One 3 lead ECG cable, Reusable SpO2(adult, paediatric, neonate) sensor, NIBP cuffs (each for Adult ,child and neonate)
15. It should have a US FDA or European CE with four digit notified body number certificate and certificate to be submitted.
16. Product should have Airworthiness RTCA DO-160 D, section 7,8,21 and Vibration standard MIL STD 810F, method 514.5 certifications

DEMO TO BE PROVIDED TO CTVS CONSULTANT AS THEY ARE LIFE SAVING EQUIPMENT

III. Terms and Conditions

1. The sealed quotation must have valid GST number
2. Rates quoted must include all types of taxes and other possible expenses. No other charges will be considered.
3. The payments will be made against delivery item with valid invoice and E-way bill through electronically viz RTGS/NEFT. The following information also to be mentioned in the quotation.
 - a. Name of the beneficiary:

- b. Account No. of the beneficiary:
- c. IFSC of the bank/branch:

d). Delivery period: Delivery period of the items shall be **10 days** from the issue of supply order. In the event of any deviation, the supply order may be cancelled.

e). The quotations will be **opened on 03/03/2025, 03:00 PM** at Opp: Mortuary Complex Block, Logistics Building No.5, 2nd Floor, Room No.2151 AIIMS Mangalagiri.

f). Vendors requested to attend at time of opening of bid.

g). Penalty: Penalty @ **0.50% per week** will levied for late delivery on supply order value of the item and will be deducted from the payment Bill. The maximum penalty will be **10 %** of total value against any of Purchase/Supply Order total value.

h). Warranty Period: Should be clearly mentioned in the quotation as per the OEM or company norms.

i). Price Quotation: Cannot exceed **Rs 5 Lakhs**.

j) Price Validity: The quoted prices should remain valid for a period of **180 days** from date of opening of NIQ.

k). The outer envelope should be super scripted as “**Supply of Transport Monitor**” at AIIMS Mangalagiri.

IV. The duly sealed and super scribed Quotations should reach **Procurement Cell, Opp: Mortuary Complex, Logistics Building No-5, Room No. 2151, AIIMS Mangalagiri, Guntur-522503 by 01:00 PM on 01/03/2025.**

Sd-
AAO (Procurement)
AIIMS, Mangalagiri