



अखिल भारतीय आयुर्विज्ञान संस्थान
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
मंगलगिरी, गुंटूर (आंध्र प्रदेश) 522503
Mangalagiri, Guntur District (Andhra Pradesh) – 522503
Website: <https://www.aiimsmangalagiri.edu.in/>

Date: **26/02/2025**

WMS/NIQ No: AIIMS/MG/Procurement/ /Vaccum Assisted Wound Closure System /

Department of Burns and Plastic Surgery.

Sub: -“NIQ for Supply of Vaccum Assisted Wound Closure System ” at AIIMS Mangalagiri.

Notice Inviting Quotations

On behalf of The Executive Director, AIIMS Mangalagiri Quotations are invited with price bid and other documents from eligible Manufacturers/ Firms/ Companies/ Authorized Agents/ Distributors/ Dealers/ Supplier Agencies for “**Supply of Vaccum Assisted Wound Closure System**” at AIIMS Mangalagiri, as per specified terms and conditions.

1. Schedule of Requirement:-

S.No	Description of Items	Required Quantity
1.	Vaccum Assisted Wound Closure System	02 No's

II. Technical Specifications

Product. Closed cyclic negative Pressure wound Therapy PU Wound Management System(VTG- 2901V2)

Accessories

1. Closed Cyclic Negative Pressure Wound Therapy PU Wound Drainage kit
- 2 Closed Cyclic Negative Pressure Wound Therapy PU Wound Drainage kit
- 3 Closed Cyclic Negative Pressure Wound Therapy PU Wound Drainage kit
- 4 Closed Cyclic Negative Pressure Wound Therapy Canister
- 5 Closed Cyclic Negative Pressure Wound Therapy Bacteria Filter

Specification

1. The system should provide controlled, Instillation, continuous or intermittent negative pressure in the range of 20-220 of Hg.
2. The system should have digital screen with battery & seal check indicator in the screen.
3. The system should have alarm to alert to user of any blockage in the tube, pressure leakage, need to change the canister or any other malfunctioning of the system.
4. The system should not generated undue noise.
5. The system should be light weight, not more than 3 kg.
6. The system should be operable in electric power supply range of 200V to 240 V
7. The system should have battery backup of at least 12-16 hrs.
8. The system should have Bluetooth enabled app for data collection of patient for wound analysis and system management.
9. The system should have CE certification.
10. The system should have safety and prevent cross infection.
11. All the under mentioned consumables / Dressings should be supplied with each system.
12. Should be US FDA or European CE with 4 digit notified body no or BIS approved for the quoted model

III. Terms and Conditions

1. The sealed quotation must have valid GST number
2. Rates quoted must include all types of taxes and other possible expenses. No other charges will be considered.

3. The payments will be made against delivery item with valid invoice and E-way bill through electronically viz RTGS/NEFT. The following information also to be mentioned in the quotation.
- a. Name of the beneficiary:
 - b. Account No. of the beneficiary:
 - c. IFSC of the bank/branch:

d). Delivery period: Delivery period of the items shall be **10 days** from the issue of supply order. In the event of any deviation, the supply order may be cancelled.

e). The quotations will be **opened on 03/03/2025, 03:00 PM** at Opp: Mortuary Complex Block, Logistics Building No.5, 2nd Floor, Room No.2151 AIIMS Mangalagiri.

f). Vendors requested to attend at time of opening of bid.

g). Penalty: Penalty @ **0.50% per week** will levied for late delivery on supply order value of the item and will be deducted from the payment Bill. The maximum penalty will be **10 %** of total value against any of Purchase/Supply Order total value.

h). Warranty Period: Should be clearly mentioned in the quotation as per the OEM or company norms.

i). Price Quotation: Cannot exceed **Rs 5 Lakhs**.

j) Price Validity: The quoted prices should remain valid for a period of **180 days** from date of opening of NIQ.

k). The outer envelope should be super scripted as “**Supply of Vacuum Assisted Wound Closure System**” at **AIIMS Mangalagiri**.

IV. The duly sealed and super scribed Quotations should reach **Procurement Cell, Opp: Mortuary Complex, Logistics Building No-5, Room No. 2151, AIIMS Mangalagiri, Guntur-522503 by 01:00 PM on 01/03/2025.**

Sd/-
AAO (Procurement)
AIIMS, Mangalagiri